



Specification of Articulation Torque Vs Angel Measuring System

Cap.	25 Nm / 50 Nm
L. Count	0.1 Nm
Max. Angel	25 Degree
Count	0.1 Degree
Accuracy	0.5% of the reading
Motor	Servo Motor
Make	Invet
HMI	Invet (For Tabular Data Only)
PC	HP 1 TB, 6 GB (For Lab. Test)
Measuring Parameter	Clock wise, counter clock wise, Break way Torque Swing Operating Angel, No. of cycle, OK Reject indication
Interface Unit	Embedded System JRAGRAU Make
Torque Sensor	JRAGRAU Make
Power	220 V AC
Testing Condition	Test will done without boot assembly
Warranty	1 Year from the date of installation

JRAGRAU INDUSTRIES

(PRESSURE, FORCE, TORQUE, DISPLACEMENT, TENSILE & SPRING TESTERS)

Mailing Address: 410, Street No. 9, Vivekanand Nagar, Bahadurgarh, Dist. Jhajjar (Haryana) – 124507.

Regd. & Factory Address: Khewat No. 37, Khata No. 55, Kila No. 33/2/2 (1-12) 9 (5-11), Main Rohtak Road,
MIE Industrial Area, Part 1, (Opp. Metro Pillar No.788) Bahadurgarh. (Haryana)

(Mob.) 9811490150, 9311990150, Email: jragrau@gmail.com